

FEB 05 2007

Amendments to the Specification:

[Para 7] Surgical tools for removing intraocular bodies are in common use. If the intraocular body is small to medium in size, its removal is not particularly problematic because the known tools are effective and generally safe to use. However, the removal of large or very large intraocular bodies is more problematic because the tools for removing small to medium sized intraocular bodies have little or no utility in removing large or very large bodies.

[Para 34] Control member 24 controls the instantaneous position of handle 18 and hence of rim 20 and net 22. Control member 24 is slideably mounted for movement along the extent of slot 18 as indicated by double-headed directional arrow 25 in Figs. 1-3. A thumb of a user easily engages control member 24 when tool[[.]]10 is held in a user's palm, supported by the fingers of the user.